

## **BLANK PAGE**



# . Scope — C

CPDC 21 (412)]

000

Ref:

CPDC 21


Instruments Subcommittee,

Dental Instruments Sectional Committee, CPDC 21; Dental Cutting and Filling

### Indian Standard

# SPECIFICATION FOR SPATULA, PLASTICS, CLEAR AND COLOURLESS, DENTAL

- 1. Scope Dimensions and other requirements for dental clear and colourless plastics spatula.
- 2. Shape and Dimensions As shown in Fig. 1.



All dimensions in millimetres.

FIG. 1 SPATULA, PLASTICS, CLEAR AND COLOURLESS, DENTAL

3. Material - Clear and colourless acrylic plastics.

### 4. Requirements

- 4.1 The spatula shall be free of cracks, scratches, pits and sharp edeges.
- 4.2 The surface of the spatula shall be sufficiently hard to withstand the abrasion while mixing silicate cements with orthophosphoric acid.
- 4.3 The spatula shall not be flexible.
- 5. Marking Mark with manufacturer's name, initials or recognized trade-mark and the country of manufacture.
- 5.1 ISI Certification Marking Details available with the Indian Standards Institution.
- 6. Packing—Wrap in wax paper or put in polyethylene bag and then pack in cartons in accordance with the trade practice or the packing may be done as agreed to between the purchaser and the supplier.

Adopted 30 January 1970

© August 1970, ISI

Price Rs 588 Gr